

ATTESTATION CE / EC CERTIFICATE

Approbation du Système Complet d'assurance Qualité / Approval of full Quality Assurance System

ANNEXE II excluant le point 4 Directive 93/42/CEE relative aux dispositifs médicaux

ANNEX II excluding section 4 Directive 93/42/EEC concerning medical devices

Pour les dispositifs de classe III, un certificat CE de conception est requis

For class III devices, a EC design certificate is required

Fabricant / Manufacturer

WiIMarc, LLC
3420 Precision Dr.,
FORT COLLINS, CO 80528 UNITED STATES

Catégorie du(des) dispositif(s) / Device(s) category

Lignes de prélèvement de gaz

Gas sample lines

Voir détails sur addendum / See attachment for additional information

GMED atteste qu'à l'examen des résultats figurant dans le rapport référencé T000429-1-R, le système d'assurance qualité - pour la conception, la production et le contrôle final - des dispositifs médicaux énumérés ci-dessus est conforme aux exigences de l'annexe II excluant le point 4 de la Directive 93/42/CEE.

GMED certifies that, on the basis of the results contained in the file referenced T000429-1-R, the quality system - for design, manufacturing, and final inspection - of medical devices listed here above complies with the requirements of the Directive 93/42/EEC, annex II excluding section 4

La validité du présent certificat est soumise à une vérification périodique ou imprévue

The validity of the certificate is subject to periodic or unexpected verification

Début de validité / Effective date : March 25th, 2019 (included)

Valable jusqu'au / Expiry date : May 3rd, 2022 (included)



Lionel DREUX
Certification Director

Identification des dispositifs / Identification of devices

Nom du dispositif médical <i>Medical device name</i>	Dénomination commerciale <i>Commercial designation</i>	Classe du DM <i>MD Class</i>
Gas Sample Line – 10ft./3m, 047" ID	003-xx ¹ -yy ² or 003-xx ¹ -yyy ²	Ila
Gas Sample Line – 10ft./3m, 042" ID	002-xx ¹ -yy ² or 002-xx ¹ -yyy ²	Ila
Gas Sample Line – 10ft./3m, 047" ID, w/ 0.8 µm Filter	003-xx ¹ F-yy ²	Ila
Gas Sample Line – 10ft./3m, 047" ID, w/ Elbow	003-xx ¹ L-yy ² or 003-xx ¹ L-yyy ²	Ila
Gas Sample Line. Accessory (Filter Only - 0.8 µm)	001-xx ¹ -yy ²	Ila

- 1) xx - Signifies connector types at each end of Gas Sample Line (e.g., MM = Male/Male, MF = Male/Female)
 2) yy or yyy - Signifies case quantities (e.g., 10, 50, or 100)

Ce certificat couvre les activités et les sites suivants :

This certificate covers the following activities and sites:

WilMarc, LLC
 3420 Precision Drive
 Fort Collins, CO, 80528 – USA

Conception, Fabrication et Contrôle final /
Design, Manufacturing and Final Control

GMED 0459



On behalf of the President
Lionel DREUX
Certification Director